

## Manufacturer's calibration requirements

It is recommended that all Royce Instruments test systems and modules be calibrated **at a minimum of every six months.**

All calibration certificates being issued by Royce Instruments will have an expiration date of six months from the date of issue. Royce Instruments will not issue certificates that exceed the six month interval for calibration.

Customers who wish to have calibration intervals longer than the recommended 6 months, are required to provide their own calibration and certification services. These calibration services should meet manufacturer's specifications and procedures.